

BACTERIA FILTRATION EFFICIENCY (BFE)

Testing performed by GAP EnviroMicrobial Services Ltd., 1020 Hargrieve Road, Unit 14, London, Ontario, Canada, N6E 1P5

A Bacterial Filtration Efficiency (BFE) test was completed according to the procedure in ASTM F2101-19 to determine the filtration efficiency of test articles by comparing the bacterial control counts upstream of the test article to the bacterial counts recovered downstream. A suspension of *S. aureus* was aerosolized using a nebulizer and delivered to the test article at a constant rate with a target delivery rate of $1.7 \times 10^3 - 3.0 \times 10^3$ colony forming units (CFU) per test article with a mean particle size of $3.0 \pm 0.3 \mu\text{m}$. The aerosolized suspension was drawn through the test article which was clamped in a six stage Andersen air sampler, at a constant flow rate of 28.3 liters per minute (LPM), for collection on bacteriological agar plates.

Challenge Microbe: *Staphylococcus aureus* ATCC 6538

Test Side: Side undetermined, flat sheet

Area Tested: ~38.5 cm²

Flow Rate: 28.3 LPM

Test Article Conditioning: 85 ± 5% RH at 25.0 ± 0.5°C for a minimum of 4 hours

Challenge Level: 5.203×10^3 CFU

Mean Particle Size: 2.95 μm

Note: A challenge level of $>3.0 \times 10^3$ CFU was accepted, as control plates remained in a countable range and particle size remained within limits for all controls.

Requirements ASTM F2100-19:

Bacteria filtration efficiency (%)

Level 1 Barrier: ≥95

Level 2 Barrier: ≥98

Level 3 Barrier: ≥98

RESULTS

Specimen #	Total CFU Recovered	Percent BFE (%)	Specimen (Pass/Fail)	FINAL RESULT
1-1	<1	>99.98	Pass	PASS Any Level
1-2	<1	>99.98	Pass	
1-3	<1	>99.98	Pass	
1-4	<1	>99.98	Pass	
1-5	<1	>99.98	Pass	

The filtration efficiency percentages were calculated using the following equation:

$$\% BFE = \frac{C - T}{C} \times 100$$

C = Challenge Level

T = Total CFU recovered downstream of test article

ANALYSIS REPORT
SCC Accreditation No.: 407

Mr. René Caiznie
 Dorena Lab

Date: August 18, 2020
 Report: 5827-001T-4B-en

IDENTIFICATION: Filter for BRP mask: SP Teflon
 Received: April 21, 2020

STANDARD:

TEST: Test Method for Filtration of 0.3 micron NaCl particles by Materials Used in Medical Face Masks GCTTG 154-202

TEST CONDITIONS: The test was subcontracted to another laboratory;
 Conditioning atmosphere: 21±5°C, 65±5% R.H. for a minimum of 4 hours;
 Air flow (L/min): 85
 Date of test: April 23 2020

RESULTS:	Individual Data					Avg.	S.D.	% CV
Resistance (mm H ₂ O):	91.8	146	123	90.1	97.3	107.0	18.6	17.4
	114	114	84.3	100	92.7			
	96.1	140	128	93.0	100			
	117	118	88.9	103	95.0			
Resistance (Pa):	680	1430	1210	684	954	1 049	183	17.4
	1120	1120	827	982	909			
	942	1480	1250	912	985			
	1150	1160	852	1010	932			
Penetration percentage (%):	0.0570	0.00900	0.0310	0.0400	0.0290	0.0320	0.0149	46.8
	0.0400	0.0220	0.0290	0.0260	0.0250			
	0.0460	0.00100	0.0300	0.0590	0.0400			
	0.0420	0.0170	0.0490	0.0320	0.0250			
Filtration efficiency (%):	99.9	100	100	100	100	100	0	0.0
	100	100	100	100	100			
	100	100	100	99.9	100			
	100	100	100	100	100			

Prepared by:


 Karine Toussaint,
 Technician

Approved by:


 Alejandro Mampone, Eng., Ph.D.
 Project Leader

Date: August 18, 2020

****For any information concerning this report, please contact Alejandro Mampone****

The reports are identified by an alphanumeric code, the letter preceding "en" refers to the version number, issued in ascending order. The electronic copy sent by CTT Group is the official report. The reported identification is based on what was observed on the received sample and/or information provided by the customer. The samples in relation to this report are retained for a period of 30 days following transmission of the report. The above reported results refer exclusively to the samples submitted for evaluation. This analysis report cannot be partly used or reproduced, unless in whole, without CTT Group prior written consent. The ISO-IEC 17025 Scope of Accreditation of CTT Group is available at www.gcttg.com. In this report, the test which number is followed by the symbol § are not covered by this accreditation. For customer's complete address, please refer to the report.

ANALYSIS REPORT
SCC Accreditation No.: 405

Mr. René Cassie
Dorma Lab

Date: August 14, 2020
 Report: 5827-001T-1A-en

IDENTIFICATION: 3D Printed Respirators with removable cartridge. Cartridge with pleated filtering media
 Received: August 10, 2020

STANDARD:

TEST: Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fluid Volume at a Known Velocity) ASTM F1862/F1862M-17

TEST CONDITIONS: Conditioning atmosphere: 21-5°C, 45-9% R.H.;
 Testing atmosphere (<1 minute): 21.5°C, 74% R.H.;
 Distance of the mask from the cartridge: 30.5cm
 Volume of fluid impacting the mask: 2.01ml
 Blood penetration detection: NAKED EYE
 Number of specimens tested: 32
 Date of test: August 12, 2020

RESULTS: Individual Data

Stream velocity (cm/s)	050				
Corresponding blood pressure (mmHg)	160				
Blood penetration (pass/fail)	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass
	Pass	Pass	Pass	Pass	Pass

Prepared by: 
 Patrick Dubois,
 Technician

Approved by: 
 Alejandro Maspons, Eng., Ph.D.
 Project Leader

Date: August 14, 2020

For any information concerning this report, please contact Alejandro Maspons.

The reports are identified by an alphanumeric code, the letter preceding "en" refers to the version number, omitted in ascending order. The electronic copy sent by CTT Group is the official report. The reported identification is based on what was observed on the received sample and/or information provided by the customer. The samples in relation to this report are retained for a period of 30 days following transmission of the report. The above reported results refer exclusively to the samples submitted for evaluation. This analysis report cannot be partly used or reproduced, unless in whole, without CTT Group prior written consent. The ISO/IEC 17025 scope of Accreditation of CTT Group is available at www.gctg.com. In this report, the text which number is followed by the symbol (s) are not covered by the accreditation. For customer's complete address, please refer to the report.



MI Integration
3100 Boulevard Industriel
Sherbrooke, Quebec J1L 1V8
Canada

KRAIBURG TPE GmbH & Co. KG
Friedrich-Schubert-Str. 2
84478 Waidhofen
Germany
US: 494-01249813021
Tel.: +49 (0) 8438 98100
Tel.: +49 (0) 8438 9810310
waidhofen@kraiburgtpe.com
www.kraiburgtpe.com
Handelsregister Traunstein
HRB 15287



Statement reg. biocompatibility testing according to "most extreme compound approach" for THERMOLAST®M

Dear Sir or Madam,

All grades, which include the same THERMOLAST® M series suffix as e.g. "MED", share the same formulation. The difference in durometer between e.g. **TM4MED** and **TM9MED**, are created by adjusting the dosage of one hardness controlling ingredient. In this case, it is a medical grade PP. The highest content of this ingredient will be found in **TM9MED**.

Consequently, the grades which have the highest durometer are considered as the "most extreme compound" of each compound series. As such, test results are representative for all grades of this compound series with lower durometers.

All tests for e.g. ISO 10993 or USP class VI conducted for **TM9MED**, yielded passing results. Therefore, the softer compound of this series, such as e.g. **TM4MED** is also covered by these test results.

Please note, Drug Master file # 22879 is assigned for the MED compound series and is valid for all hardness ranges, including **TM4MED**.

Thus, we are convinced our material, **TM4MED** will pass further biocompatibility tests, including but not limited to: ISO 10993-5 (passed, up to date), 10993-10 and 10993-11.

Medizinprodukte Berater
Oliver Kluge
Market Segment Manager Medical Applications
KRAIBURG TPE GmbH & Co. KG



PDG
KRAIBURG TPE Verwaltungs GmbH
Str. Waidhofen
Handelsregister Traunstein
HRB 17274
GF: Franz Holtenacker

Licence Number **14230** Numéro de la licence

Medical Device
Establishment Licence

Licence d'établissement
pour les instruments médicaux

DORMA DBA DORMA SOINS DU SOMMEIL

1055 BEAVER HALL HILL
 SUITE 400
 MONTREAL, QUEBEC
 CANADA
 H2Z 1S5

This licence is issued in accordance with the Medical Devices Regulations of the Food and Drugs Act for the following activities:

Cette licence est délivrée conformément à la Loi sur les aliments et drogues, règlement sur les instruments médicaux pour les activités qui suivent:

	Distributor / Distributeur	Importer / Importateur	Manufacture Devices for Distribution / Fabricant d'instruments médicaux pour distribution
Class I / Classe I	No / Non	Yes / Oui	Yes / Oui
Class II / Classe II	No / Non	Yes / Oui	
Class III / Classe III	No / Non	Yes / Oui	
Class IV / Classe IV	No / Non	No / Non	

Attestation made :


Attestations faites :

The establishment has documented procedures in place in respect of: <ul style="list-style-type: none"> distribution records complaint handling recalls mandatory problem reporting handling, storage, delivery installation corrective action servicing 	[Y] [Y] [Y] [Y] [Y] [Y] [Y] [Y]	L'établissement a mis en oeuvre une procédure écrite concernant: <ul style="list-style-type: none"> les registres de distribution les plaintes les rappels rapports d'incident obligatoires la manutention, le stockage, la livraison l'installation, les mesures correctives l'entretien
---	--	---

Site listing begins on the back of this page

Liste des sites commence au verso de cette page

Issue Date, date de délivrance: 2020-06-30

Minister of Health Ministre de la santé	Countersigned: Director, Medical Devices Compliance Program or delegated authority Contresigné par: Directeur, Programme de la conformité des matériels médicaux ou autorité déléguée  Anik Michelle Chartrand
--	---

This licence is the property of the Medical Devices Compliance Program and must be returned upon demand.
 Cette licence appartient au Programme de la conformité des matériels médicaux et doit être retournée sur demande.